

## MANUFACTURER'S NOTIFICATION

### CE mark removal from labels and packaging of Socorex Qualitix™ pipette tips Information restricted to Swiss market

Socorex, a manufacturer of high-quality liquid handling instruments and worldwide supplier of laboratory products, is committed to continuously improving its products and procedures.

Following changes in the CE IVD regulation in 2022, Socorex removed the CE IVD marking on packaging and labels of all new productions of instruments and Qualitix™ pipette tips, and updated its relevant documentation.


Regarding stock of products manufactured under the former regulation, Socorex relabeled all commercial packaging by removing the CE marking where possible, and otherwise, covered the CE information with a sticker, to prevent any confusion related to CE conformity.

To maintain the aseptic properties of sterile products, the sticker is only added on the outer product packaging.

At this stage, it is important to emphasize that all instruments and consumables meet the current regulation requirements. This change has no impact on our product quality, meeting the highest standards and excellent reliability.

Should you have any question regarding this matter, please contact us.


**SOCOREX ISBA SA**



Julien Ruchet  
Chief Executive Officer, CEO

[julien.ruchet@socorex.com](mailto:julien.ruchet@socorex.com)

**SOCOREX ISBA SA**



Jean-Marc Ammann  
Quality System & Regulatory Affairs Manager

[jeanmarc.ammann@socorex.com](mailto:jeanmarc.ammann@socorex.com)

Ecublens, July 9<sup>th</sup>, 2024